BioRegen
Independent Testing, Trusted Results

Report ID: [redacted]-ENDO-03072026-306

Validation Key: [redacted]

contact@bioregen.com
www.bioregen.com

Limulus Amebocyte Lysate (LAL) – Report of Analysis

Test Performed	Endosafe - PTS LAL Assay
Test Date	03/07/2026
Analyst	VICTOR
Instrument	Endosafe® nexgen-PTS™ (SN: 25201216)

Sample Information

Company Name	[redacted]
Sample Tested	AOD-9604
Sample Lot Number	SYN-AOD5-0002
Sample Amount	5mg
Vials/Units Received	1

Reagent Information

Cartridge Type	Endotoxin- LAL (Charles River)
Cartridge Lot #	2523801115
Cartridge Calibration Code	512041273424
Detection Range	0.05 EU/mL – 5.0 EU/mL
Spike Value	0.428 EU/mL
Spike Recovery	79% (Pass)
Test Suitability	Pass

Endotoxin Quantitation Results

Sample ID	Replicates	Measured EU/mL	Specification	Pass/Fail
AOD-9604	1, 2	< 0.05 EU/mL	≤ 0.50 EU/mL	Pass

Interpretation:
Endotoxin level in **AOD-9604** was measured using chromogenic LAL assay with detection at 405 nm. The result of **< 0.05 EU/mL** indicates that the endotoxin content is **below** the acceptable threshold of 0.5 EU/mL. Sample **PASSES** the endotoxin limit test.

Analyst Signature: _____
Date: 03/07/2026

To verify the authenticity of this report, please contact **BioRegen** at **contact@bioregen.com** or reference the **Report ID** and **Validation Key** provided above.

Alternatively, **scan the QR code** below to access the official record on our website.

[redacted]